

TRANSLATION. ORIGINAL TEXT IN FINNISH.

**Exemption from safety licence for the use of equipment that generates ionising radiation electrically when the equipment operates at a maximum voltage of 30 kV**

By virtue of Section 17, paragraph 1, point 3 of the Radiation Act (592/1991), the Radiation and Nuclear Safety Authority has decided that a safety licence is not required for the use of equipment that generates ionising radiation electrically if the equipment operates at a maximum voltage of 30 kV and it does not generate a higher dose rate than 1  $\mu\text{Sv/h}$  at any point within 10 cm of the device's accessible surfaces.

This decision concerns the use, manufacture and trade of such equipment pursuant to Section 11, paragraph 2 of the Radiation Act and the activities related to this.

Director General

Petteri Tiippana

Director

Eero Kettunen